




D.H.H.srl
SHIELD
MEDICAL SOLUTIONS



LA SOCIETÀ D.H.H. SRL NASCE NEL 2017 PER LA PRODUZIONE E LA VENDITA DI SISTEMI PER LA PREVENZIONE INDIVIDUALE. NELLO SPECIFICO CI OCCUPIAMO DI DISPOSITIVI DI PROTEZIONE INDIVIDUALE (MATERIALI DPI). IL NOSTRO CORE BUSINESS COMPRENDE LA PRODUZIONE DI MASCHERINE 3PLY PERSONALIZZABILI, 3PLY TIPO IIR MEDICHE CHIRURGICHE, FFP2 E FFP3.

D.H.H. LTD. IS A COMPANY, BORN IN 2017, BASED ON THE PRODUCTION AND SELLING OF PERSONAL PROTECTIVE EQUIPMENT (PPE). OUR CORE BUSINESS IS BASED ON THE PRODUCTION OF DIFFERENT TYPES OF FACE MASKS; WE PRODUCE CUSTOMISED 3PLY MASKS, 3PLY IIR SURGICAL MASKS, FFP2 AND FFP3 MASKS.





LE NOSTRE STRUTTURE SONO LOCALIZZATE IN DIVERSI PAESI DEL MONDO TRA CUI STATI UNITI, CINA E AUSTRALIA. IN ITALIA ATTUALMENTE LA NOSTRA SEDE OPERATIVA È SITUATA NELLA CITTÀ DI POMEZIA, IN PROVINCIA DI ROMA.

OUR FACILITIES ARE LOCATED IN SEVERAL COUNTRIES WORLDWIDE INCLUDING THE UNITED STATES, CHINA AND AUSTRALIA. IN ITALY, OUR OPERATIONAL HEADQUARTERS ARE CURRENTLY LOCATED IN THE CITY OF POMEZIA, IN THE PROVINCE OF ROME.



INNOVATIVE TECHNOLOGIES



LA NOSTRA MISSIONE È DI PRODURRE MASCHERINE DI ALTA QUALITÀ A LIVELLO GLOBALE E LOCALE, COMBINANDO LE NOSTRE TECNOLOGIE INNOVATIVE LEADER A LIVELLO MONDIALE E IL FOCUS SULLA QUALITÀ DEI NOSTRI PRODOTTI APPREZZATI DA CLIENTI IN TUTTO IL MONDO.

OUR MISSION IS TO PRODUCE HIGH QUALITY MASKS GLOBALLY AND LOCALLY, COMBINING OUR LEADING INNOVATIVE TECHNOLOGIES AND THE FOCUS ON THE QUALITY OF OUR PRODUCTS APPRECIATED BY CUSTOMERS WORLDWIDE.



STERILE CLEANROOM

ALL'INTERNO DELLA NOSTRA FABBRICA ABBIAMO INSTALLATO UNA CAMERA BIANCA ISO CLASSE 5:100000 PARTICELLE 0,1 μ M, UN AMBIENTE LA CUI CARATTERISTICA PRINCIPALE È LA PRESENZA DI UN BASSISSIMO CONTENUTO DI MICRO PARTICELLE DI POLVERE IN SOSPENSIONE, CHE CONSENTE LA PRODUZIONE DI GRANDI QUANTITÀ DI MASCHERINE IN MANIERA COMPLETAMENTE STERILE.


INSIDE OUR FACTORY WE HAVE INSTALLED A CLEAN ROOM ISO CLASS 5:100000 PARTICLES 0.1 μ M, AN ENVIRONMENT CHARACTERISED BY A VERY LOW CONTENT OF SUSPENDED MICRO DUST PARTICLES, WHICH ALLOWS THE PRODUCTION OF LARGE QUANTITIES OF MASKS IN A COMPLETELY STERILE WAY.





LE MASCHERINE MEDICHE CHIRURGICHE IIR SONO IPOALLERGENICHE, HANNO UN'OTTIMA VESTIBILITÀ E PRESENTANO UNA BUONA RESISTENZA ALLA PENETRAZIONE BATTERICA, GRAZIE AL SUO POTERE FILTRANTE $BFE \geq 99\%$.
MASCHERINE CONFORMI ALLO STANDARD EUROPEO EN 14683:2019.

IIR MEDICAL SURGICAL MASKS ARE HYPOALLERGENIC, PROVIDE AN EXCELLENT FIT AND HAVE GOOD RESISTANCE TO BACTERIAL PENETRATION DUE TO ITS $BFE \geq 99\%$ FILTERING POWER. OUR MASKS COMPLY WITH THE EUROPEAN STANDARD EN 14683:2019.





MASCHERINA CHIRURGICA 3PLY TIPO IIR

TESSUTO NON TESSUTO
PROTETTIVO

PROTECTIVE
NONWOVEN FABRIC

TESSUTO FUSO SOFFIATO
FILTRANTE AL 99%

99% FILTRATION
MELTBLOWN FABRIC

TESSUTO NON TESSUTO
DERMOCOMPATIBILE


SKIN FRIENDLY
NONWOVEN FABRIC

THREE LAYER FILTRATION DESIGN

ULTRASONIC WELDING DOTS

METAL CORE PLASTIC
NOSEBRIDGE

ELASTIC EARLOOP



FILTRAZIONE PROTETTIVA
A TRE STRATI


PUNTI DI SALDATURA
AD ULTRASUONI

NASELLO IN PLASTICA
CON ANIMA IN METALLO

LACCI ELASTICI
PER ORECCHIE

La mascherina Chirurgica 3-ply tipe IIR proteggere noi e gli altri essendo un dispositivo medico a 3 veli in TNT (tessuto non tessuto) melt-brown ad alta nitratura, con strato esterno antispruzzo e strato interno dermocompatibile. Le mascherine vengono prodotte in ambiente sterile, in una camera bianca, per una maggiore sterilizzazione.

Questo dispositivo medico può rappresentare anche un'opportunità per promuovere un brand o un prodotto. Un gadget che ha un'altissima visibilità poiché le scritte o i loghi presenti sulle mascherine vengono notati da chiunque entri in contatto con le persone che la indossano.



DIMENSIONE: 92mm X 177mm

VALIDITÀ: 2 ANNI DALLA DATA DI PRODUZIONE

MATERIALI

TESSUTO NON TESSUTO PROTETTIVO
FILTRAGGIO CON TESSUTO MEL TBLOWN FINO AL 99%
TESSUTO NON TESSUTO DERMOCOMPATIBILE

STANDARD: IIR


FILTRAGGIO: FINO AL 99% DEI MICRORGANISMI

PRODUCT DESCRIPTION

The IIR medical surgical masks is designed to protect the wearer and other people.

Our 3ply mask is made of three layers including an outer hydrophobic non-woven layer, a middle melt-blown layer, and an inner soft absorbent non-woven layer. The masks are produced in a sterile environment, a sterile clean room, for greater sterilisation.

Our 3-ply customised masks represent a great opportunity for you to promote your business. With our masks you won't go unnoticed anymore!



PRODUCT SIZE: 92mm X 177mm

EXPIRATION DATE: WITHIN 2 YEARS FROM THE PRODUCTION DATE

MATERIALS

PROTECTIVE. NON-WOVEN FABRIC
99% FILTRATION THANKS TO A MELT-BLOWN FABRIC
DERMOCOMPATIBLE NON-WOVEN FABRIC


STANDARD: TYPE IIR

FILTRATION EFFICIENCY: CATCH 99% OF MICROORGANISMS




CONFEZIONE

PACKAGING



**MASCHERINA
CHIRURGICA IIR 2
MONOUSO
CE 2163**





ATTESTATION OF CONFORMITY

Certificate Nr: MDD-284

In conformity to the European Economic Commission 93/42/EEC Medical Devices Directive on harmonisation of laws, regulations and administrative documentation of Member States on Medical Devices and European Commission directive 2007/47/EC amending Medical Devices Directive dated 05 September 2007,

the products manufactured by

D.H.H. SRL

at the following address

Via Campobello 30, 00071 Pomezia (RM), ITALY

EN 14683:2019+AC:2019 Medical Face Masks

Model: DHH20

Type IIR

are tested according to the following initial type tests by the manufacturer

Technical standard EN 14683:2019+AC:2019 Medical face masks - Requirements and test methods

For the assessment of conformity, the following documents were also applied to:
Results of laboratory tests Çevre Industrial Testing Laboratory Bacterial Filtration Efficiency, Microbial Cleanliness, Differential Pressure and Splash Resistance Pressure tests.

UNIVERSAL CERTIFICATION has evaluated production, design, intended use, risk evaluation according to safety purpose, product itself and add-on components (if exists) and product technical drawings of the medical face masks manufactured and designed for use during the medical operations or similar medical situations with same requirements which require restriction of infectious materials to be spread to patients. With this certificate, it is approved that the product fulfills all essential requirements and the related rules of 93/42/EEC Medical Devices Directive (MDD) Class I are applied. The information on the packaging for the above listed products covers the necessary information stated in Annex I, §13, of the Medical Devices Directive (93/42/EEC) or Annex I, §23, of the Medical Device Regulation (EU) 2017/745. This information includes; reference to EN 14683 standard, type of mask (as indicated in Table 1) and other relevant information given in EN ISO 15223-1:2016 and EN 1041:2008+A1:2013. It is considered to be suitable to attach a CE mark, as seen below, on your products in accordance with the information given in this certificate with publishing an EU Declaration of Conformity.

This certificate is issued on 19/10/2020 and valid until 18/10/2021 with the conditions that no change has been made with the product references and no change in the production process or not suspended or withdrawn for any reason.

ISTANBUL - 19/10/2020



Verify the validity with the QR Code

Suat KACMAZ
UNIVERSAL CERTIFICATION
Director

This certificate will be in the absence of any changes in standard and legal terms, and with the surveillance audits to be conducted annually following the surveillance audits, updating the publication date without changing the certificate number.



TECHNICAL EVALUATION REPORT

REPORT DATE / NO: 19.10.2020 / 10-2020-T0460

Manufacturer: D.H.H. SRL

Address: Via Campobello 30, 00071 Pomezia (RM), ITALY

The medical masks manufactured by the above manufacturer, are evaluated based on the Annex ZA of harmonised standard EN 14683/AC:2019 and the essential health and Safety requirements of 93/42/EEC, Medical Device Directive for Class I products on a voluntary basis upon the manufacturer request.

Product Description: Medical Face Mask

Model: DHH20



As a third party evaluation, the technical file provided by the manufacturer is evaluated and the samples provided by the manufacturer are tested according to Annex ZA of the EN 14683/AC:2019 standard. See Annex I: Test report provided by Çevre Endüstriyel Analiz Laboratuvarı 16.10.2020 2022987E-R1 date and with report number.

This report or the issued certificate, in case the report is positive, does not take over or change the sole responsibility of the manufacturer covered under 93/42/EEC Medical Device Directive. The manufacturer shall fulfil all responsibilities for Class I products under 93/42/EEC Medical Device Directive.

UFR-383 12.12.2018 Rev.01

Page 1/3



The results of the evaluation are as follows;

A- Review of the technical file

The manufacturer owns a technical file based on the requirements of 93/42/EEC Medical Device Directive in which the essential health and Safety requirements for Class I products are handled and have documented procedures to fulfil these requirements. The positive result of this report or the possible certificate to be issued based on positive result of this report shall not be used as the share of the responsibility of manufacturer on the fulfilment of any responsibility to be fulfilled before putting the product on the EU market.

B- Product Test Results

The tests referenced in Annex ZA are conducted on the samples provided by the manufacturer and the results are evaluated;

1. Biocompatibility

In the evaluation of the technical file, it was observed that the manufacturer has established a mechanism for the evaluation of raw materials or semi-finished goods or their biocompatibility. The manufacturer claims that the request and evaluation of proof for biocompatibility of the goods is an essential part of the procurement policy and declares that the produced masks are compliant with the biocompatibility requirements and have authorised responsible staff members for ensuring the success of this policy. It is considered that the manufacturer have an effective policy for the biocompatibility of the product.

2. Bacteria Filtration Efficiency

At least 5 samples are subjected to a bacteria aerosol with a flow rate of 28.3 L/min for 1 minutes with a test setup defined in the Annex B of EN 14683/AC:2019 standard. With the results of the incubation of samples taken in different particle sizes are shown in the annexed test report.

The minimum bacteria filtration efficiency performance required by each performance classes are shown below;

Test	Type I*	Type II	Type IIR
Bacterial Filtration Efficiency (BFE), (%)	≥ 95	≥ 98	≥ 98

* Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

According to the evaluation of the results of 5 samples tested, the minimum bacteria filtration efficiency is given as 99,6%. According to this result, the bacteria filtration efficiency performance of the masks is classified as Type IIR.

It was observed that the average positive control values and negative control value is also reported as a confidence parameter of the test result are meaningful.

3. Microbial Cleanliness (Bioburden)

It is expected to have the number of colony forming units per gram to be lower than 30 for all performance class of masks according to the test result based on ISO 11737-1 standard. In the evaluation of the test result, the maximum count of the colony forming unit is reported as 3. For this test result the samples complies the requirement for all performance classes (Type I, Type II and Type IIR).

UFR-383 12.12.2018 Rev.01

Page 2/3



4. Differential Pressure

The test is conducted to measure the breathing resistance as the differential pressure and the expected result for Type I and Type II classes is not to be higher than 40 Pa/cm² and for Type IIR class not to be higher than 60 Pa/cm².

According to the test results, the highest differential pressure measured is 34,6 Pa/cm² and the samples complies the requirement for all performance classes (Type I, Type II and Type IIR).

5. Splash Resistance Pressure

In the test, done according to ISO 22609:2004 the product's splash resistance is expected to be equal or higher than 16kPa for the Type 2R class.

All 15 samples tested were able to provide Type IIR performances as 16kPa resistance.

C- Summary and Conclusion

Evaluation	Requirement	Result	Classification
Bacterial Filtration Efficiency (BFE), (%)	≥ 95 % - Type I ≥ 98 % - Type II ≥ 98 % - Type IIR	99,6 %	Type I Type II Type IIR
Differential pressure (Pa/cm ²)	< 40 - Type I < 40 - Type II < 60 - Type IIR	34,6	Type I Type II Type IIR
Splash resistance pressure (kPa)	Not Required - Type I Not Required - Type II ≥ 16 - Type IIR	> 16	Type IIR
Microbial cleanliness (cfu/g)	≤ 30 - Type I ≤ 30 - Type II ≤ 30 - Type IIR	3	Type I Type II Type IIR
Overall Performance Classification			Type IIR

- End of Report -

UNIVERSAL CERTIFICATION
VE GÖZETİM HİZMETLERİ
TİC. LTD. ŞTİ.
Hicpazı Bulvarı, Kat: 5, 06450
Yükseköğretim Kurumları
Tel: 0312 455 80 80 Faks: 0312 455 80 06
Sertifika No: 00135473
SUA KACMAZ
UNIVERSAL CERTIFICATION
Director

UFR-383 12.12.2018 Rev.01

Page 3/3



MASCHERINE FACCIALI FILTRANTI FFP2 NR

FFP2 NR FILTERING FACE MASKS



LE MASCHERINE FFP2 NR SONO IPOALLERGENICHE E IMPERMEABILI, HANNO UN'OTTIMA VESTIBILITÀ, GARANTISCONO UN AGEVOLE FLUSSO D'ARIA IN ENTRAMBE LE DIREZIONI, POICHÉ TRASPIRANTI, E PRESENTANO UNA BUONA RESISTENZA ALLA PENETRAZIONE BATTERICA, GRAZIE AL SUO POTERE FILTRANTE BFE $\geq 99\%$. MASCHERINE CONFORMI ALLO STANDARD EUROPEO EN 149:2001

FFP2 NR FILTERING MASKS ARE HYPOALLERGENIC AND WATERPROOF, THEY PROVIDE AN EXCELLENT FIT, GUARANTEE AN EASY AIR FLOW IN BOTH DIRECTIONS AND HAVE A GOOD RESISTANCE TO BACTERIAL PENETRATION, THANKS TO ITS BFE $\geq 99\%$ FILTERING POWER.
OUR MASKS COMPLY WITH THE EUROPEAN STANDARD EN 149:2001



MASCHERINE FACCIALI FILTRANTI FFP2 NR

FFP2 NR FILTERING FACE MASKS

1.
PROTECTIVE
NONWOVEN 50 GSM

2.
HOT AIR
FABRIC 50 GSM

3.
MELT-BLOWN
25 GSM

4.
SKIN FRIENDLY
NONWOVEN 25 GSM



IL PRIMO STRATO ESTERNO
È IN TESSUTO NON TESSUTO


IL SECONDO È UNO STRATO
DI COTONE SOFFIATO

IL TERZO STRATO È COMPOSTO
DA UN FILTRO MELT-BLOWN

IL TERZO STRATO È COMPOSTO
DA UN FILTRO MELT-BLOWN

IL QUINTO STRATO INTERNO È UN
MATERIALE DERMOCOMPATIBILE

PRODUCT PARAMETERS



Tutte le misure vengono rilevate a mano. Potrebbero esserci degli errori.
Si prega di fare riferimento ai parametri presenti nel prodotto.


All dimensions are measured by hand. There may be errors.
Please refer to the actual products.

ULTRASONIC WELDING DOTS

SKIN FRIENDLY
NONWOVEN

METAL CORE PLASTIC
NOSEBRIDGE

ELASTIC EARLOOP



PUNTI DI SALDATURA
AD ULTRASUONI

TNT
DERMOCOMPATIBILE

NASELLO IN PLASTICA
CON ANIMA IN METALLO

LACCI ELASTICI
PER ORECCHIE



CONFEZIONE

PACKAGING



UNIVERSAL

Verify the validity with the QR code

**EU TYPE EXAMINATION CERTIFICATE****Certificate No: 2163-PPE-1587**

Respiratory protective devices, filtering half masks to protect against particles manufactured by

D.H.H. SRL

Via Campobello 30, 00071 Pomezia (RM) Sede ITALY

are tested and evaluated according to

**EN 149:2001 + A1:2009 Respiratory Protective Devices -
Filtering Half Masks to Protect Against Particles -
Requirements, Testing, Marking**

Based on the type examination conducted with the evaluation of test reports, technical file according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 5, it is approved that the product meets the requirements of the regulation.

Product Definition**Model:** DHH20M

Filtering half mask

Classification: FFP2 NR

Here by the manufacturer is allowed to use notified body number (2163) and can fix CE mark, as shown below, on the Category III product models given above, with;

- Issuing an appropriate EU Declaration of Conformity according to **Personal Protective Equipment Regulation (EU) 2016/425 Annex 9**.
- Ongoing successful performance in fulfilment of the requirements set out in Personal Protective Equipment Regulation (EU) 2016/425 and harmonised standards, ensured by assessments based on **Annex 7 (Module C2) or Annex 8 (Module D)** of the regulation no later than 1 year from the beginning of serial production

This certificate is initially issued on **19/10/2020** and will be valid for 5 years, if there is no change in the relevant harmonised standard affecting the essential health and safety requirements.**Suat KACMAZ**
UNIVERSAL CERTIFICATION
Director



MASCHERINE PERSONALIZZATE

CUSTOMISED MASKS





OPPORTUNITÀ PER I BRAND SI ESSERE VISIBILI

A REAL OPPORTUNITY FOR COMPANIES TO BE VISIBLE



CON LA NOSTRA
TECNOLOGIA PUOI
POSIZIONARE LA
STAMPA ESATTAMENTE
AL CENTRO

THANKS TO OUR HIGH
TECHNOLOGY,
YOU CAN PLACE
THE PRINT
EXACTLY IN THE
CENTRE

CUSTOMISED MASKS



☀ TESTATO E CERTIFICATO UNIVERSAL

☀ TESTED AND CERTIFIED BY UNIVERSAL

☀ STANDARD MEDICO ELEVATO TIPO IIR

☀ HIGHEST MEDICAL STANDARDS: TYPE IIR

☀ 99.8% EFFICIENZA FILTRAGGIO

☀ 99.8 % OF FILTRATION EFFICIENCY

☀ 100% UTILIZZO SICURO, LA STAMPA NON RILASCIAM RESIDUI O ODORI

☀ 100% SAFE. NO PRINT RESIDUALS OR ODOURS MICRODROPLETS

MICRO GOCCE



MICRODROPLETS

FUMO



SMOG

ODORI



ODOURS

POLVERE



DUST

SPORT CLUB FACE MASKS



STILE E SICUREZZA!

Le nostre mascherine non sono solo estremamente eleganti e personalizzabili, ma anche sicure al 100%.

Tutte le nostre mascherine sono sterili e costituite da tre strati di tessuto non tessuto ad alta tecnologia di filtrazione.

STYLE AND SAFETY!


Not only our masks are extremely stylish and customisable, but they are also 100% safe! All our masks are sterile and made of three-layers of non-woven fabric with a high filtration technology.



NATIONAL FLAGS



Una vera
innovazione nel
settore turistico! Le
nostre mascherine a
3 strati
personalizzate
saranno un perfetto
souvenir da portare
a casa!




A real innovation in
the tourism sector!

Our
customised 3-ply
masks will be the
perfect
souvenir to bring
back home!

KIDS MASKS









ATTESTATION OF CONFORMITY

Certificate Nr: MDD-284

In conformity to the European Economic Commission 93/42/EEC Medical Devices Directive on harmonisation of laws, regulations and administrative documentation of Member States on Medical Devices and European Commission directive 2007/47/EC amending Medical Devices Directive dated 05 September 2007,

the products manufactured by

D.H.H. SRL

at the following address

Via Campobello 30, 00071 Pomezia (RM), ITALY

EN 14683:2019+AC:2019 Medical Face Masks

Model: DHH20

Type IIR

are tested according to the following initial type tests by the manufacturer

Technical standard EN 14683:2019+AC:2019 Medical face masks - Requirements and test methods

For the assessment of conformity, the following documents were also applied to:
Results of laboratory tests Çevre Industrial Testing Laboratory Bacterial Filtration Efficiency, Microbial Cleanliness, Differential Pressure and Splash Resistance Pressure tests.

UNIVERSAL CERTIFICATION has evaluated production, design, intended use, risk evaluation according to safety purpose, product itself and add-on components (if exists) and product technical drawings of the medical face masks manufactured and designed for use during the medical operations or similar medical situations with same requirements which require restriction of infectious materials to be spread to patients. With this certificate, it is approved that the product fulfills all essential requirements and the related rules of 93/42/EEC Medical Devices Directive (MDD) Class I are applied. The information on the packaging for the above listed products covers the necessary information stated in Annex I, §13, of the Medical Devices Directive (93/42/EEC) or Annex I, §23, of the Medical Device Regulation (EU) 2017/745. This information includes; reference to EN 14683 standard, type of mask (as indicated in Table 1) and other relevant information given in EN ISO 15223-1:2016 and EN 1041:2008+A1:2013. It is considered to be suitable to attach a CE mark, as seen below, on your products in accordance with the information given in this certificate with publishing an EU Declaration of Conformity.

This certificate is issued on 19/10/2020 and valid until 18/10/2021 with the conditions that no change has been made with the product references and no change in the production process or not suspended or withdrawn for any reason.

ISTANBUL - 19/10/2020



Verify the validity with the QR Code

Suat KACMAZ
UNIVERSAL CERTIFICATION
Director

This certificate will be in the absence of any changes in standard and legal terms, and with the surveillance audits to be conducted annually following the surveillance audits, updating the publication date without changing the certificate number.



TECHNICAL EVALUATION REPORT

REPORT DATE / NO: 19.10.2020 / 10-2020-T0460

Manufacturer: D.H.H. SRL

Address: Via Campobello 30, 00071 Pomezia (RM), ITALY

The medical masks manufactured by the above manufacturer, are evaluated based on the Annex ZA of harmonised standard EN 14683/AC:2019 and the essential health and Safety requirements of 93/42/EEC, Medical Device Directive for Class I products on a voluntary basis upon the manufacturer request.

Product Description: Medical Face Mask

Model: DHH20



As a third party evaluation, the technical file provided by the manufacturer is evaluated and the samples provided by the manufacturer are tested according to Annex ZA of the EN 14683/AC:2019 standard. See Annex I: Test report provided by Çevre Endüstriyel Analiz Laboratuvarı 16.10.2020 2022987E-R1 date and with report number.

This report or the issued certificate, in case the report is positive, does not take over or change the sole responsibility of the manufacturer covered under 93/42/EEC Medical Device Directive. The manufacturer shall fulfil all responsibilities for Class I products under 93/42/EEC Medical Device Directive.

UFR-383 12.12.2018 Rev:01

Page 113



The results of the evaluation are as follows;

A- Review of the technical file

The manufacturer owns a technical file based on the requirements of 93/42/EEC Medical Device Directive in which the essential health and Safety requirements for Class I products are handled and have documented procedures to fulfil these requirements. The positive result of this report or the possible certificate to be issued based on positive result of this report shall not be used as the share of the responsibility of manufacturer on the fulfilment of any responsibility to be fulfilled before putting the product on the EU market.

B- Product Test Results

The tests referenced in Annex ZA are conducted on the samples provided by the manufacturer and the results are evaluated;

1. Biocompatibility

In the evaluation of the technical file, it was observed that the manufacturer has established a mechanism for the evaluation of raw materials or semi-finished goods or their biocompatibility. The manufacturer claims that the request and evaluation of proof for biocompatibility of the goods is an essential part of the procurement policy and declares that the produced masks are compliant with the biocompatibility requirements and have authorised responsible staff members for ensuring the success of this policy. It is considered that the manufacturer have an effective policy for the biocompatibility of the product.

2. Bacteria Filtration Efficiency

At least 5 samples are subjected to a bacteria aerosol with a flow rate of 28.3 L/min for 1 minutes with a test setup defined in the Annex B of EN 14683/AC:2019 standard. With the results of the incubation of samples taken in different particle sizes are shown in the annexed test report.

The minimum bacteria filtration efficiency performance required by each performance classes are shown below;

Test	Type I*	Type II	Type IIR
Bacterial Filtration Efficiency (BFE), (%)	≥ 95	≥ 98	≥ 98

* Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

According to the evaluation of the results of 5 samples tested, the minimum bacteria filtration efficiency is given as 99,6%. According to this result, the bacteria filtration efficiency performance of the masks is classified as Type IIR.

It was observed that the average positive control values and negative control value is also reported as a confidence parameter of the test result are meaningful.

3. Microbial Cleanliness (Bioburden)

It is expected to have the number of colony forming units per gram to be lower than 30 for all performance class of masks according to the test result based on ISO 11737-1 standard. In the evaluation of the test result, the maximum count of the colony forming unit is reported as 3. For this test result the samples complies the requirement for all performance classes (Type I, Type II and Type IIR).

UFR-383 12.12.2018 Rev:01

Page 213



4. Differential Pressure

The test is conducted to measure the breathing resistance as the differential pressure and the expected result for Type I and Type II classes is not to be higher than 40 Pa/cm² and for Type IIR class not to be higher than 60 Pa/cm².

According to the test results, the highest differential pressure measured is 34,6 Pa/cm² and the samples complies the requirement for all performance classes (Type I, Type II and Type IIR).

5. Splash Resistance Pressure

In the test, done according to ISO 22609:2004 the product's splash resistance is expected to be equal or higher than 16kPa for the Type 2R class.

All 15 samples tested were able to provide Type IIR performances as 16kPa resistance.

C- Summary and Conclusion

Evaluation	Requirement	Result	Classification
Bacterial Filtration Efficiency (BFE), (%)	≥ 95 % - Type I ≥ 98 % - Type II ≥ 98 % - Type IIR	99,6 %	Type I Type II Type IIR
Differential pressure (Pa/cm ²)	< 40 - Type I < 40 - Type II < 60 - Type IIR	34,6	Type I Type II Type IIR
Splash resistance pressure (kPa)	Not Required - Type I Not Required - Type II ≥ 16 - Type IIR	> 16	Type IIR
Microbial cleanliness (cfu/g)	≤ 30 - Type I ≤ 30 - Type II ≤ 30 - Type IIR	3	Type I Type II Type IIR
Overall Performance Classification			Type IIR

- End of Report -

UNIVERSAL CERTIFICATION
VE GÖZETİM HİZMETLERİ
TİC. LTD. ŞTİ.
Hacıoğlu Mahallesi, Pınarbaşı, 32. Blok, No:4434 Y. Duhulu - Çanakkale - İSTANBUL, T: +90 216 455 80 80 F: +90 216 455 80 80
E: info@universalcert.com.tr
Sertifika No: 2022987E-R1
SUA KACMAZ
UNIVERSAL CERTIFICATION
Director

UFR-383 12.12.2018 Rev:01

Page 313



D.H.H. srl

WWW.SHIELDMEDICALSOLUTIONS.IT - Tel. 06 86 29 98 79

D.H.H. srl

via Campobello 30, 00040 Pomezia (RM), Italy.

+39 06 86299879

www.shieldmedicalsolutions.it

Mail: info@shieldmedicalsolutions.it